


Optical Density Scan



Semrock Part Number: **FF01-450/70**

Lot Number: **113043**

Note: measurement taken at 45° incidence, average polarization.
measurement accuracy is limited to about OD = 5.

The graph above is representative of the coating lot from which your filter was manufactured. The spectrum of your actual filter(s) might differ slightly within manufacturing tolerances and is certified to meet all optical specifications as detailed in the Semrock online catalog or in the associated custom specification drawing.